



Fogg System Company, Inc.
 15592 East Batavia Drive
 Aurora, Colorado 80011
 United States of America

Toll Free: 1 800 525-0292
 Telephone: 303 344-1883
 FAX: 303 344-1780

ECG WAVEFORM MEMORY MODULE INFORMATION

Part Number 7147-0060
 PQRST & PQRS with ST Elevation and Depression

ABBREV.	DESCRIPTION	UNIT	SHAPE			
			AN	BN	AW	BW
Pa	P wave amplitude	μV	200	200	200	200
Pd	P wave duration	ms	80	80	80	80
PRi	PR interval	ms	120	130	130	130
Qa	Q wave amplitude	μV	—	—	—	—
Qd	Q wave duration	ms	—	—	—	—
Ra	R wave amplitude	μV	1000	1000	1000	1000
Rd	R wave duration	ms	—	—	—	—
Rsl	R wave leading slope	mV/s	—	—	—	—
Sa	S wave amplitude	μV	—	—	—	—
Sd	S wave duration	ms	—	—	—	—
QRSd	QRS duration	ms	100	100	100	100
ST-Ja	ST amplitude at J	μV	0	0	300	-300
ST-Ea	ST amplitude at end	μV	0	0	300	-300
STd	ST duration	ms	80	50	90	90
STsl	ST slope	mV/s	—	—	—	—
Ta	T wave amplitude	μV	400	400	0	0
Td	T wave duration	ms	60	80	0	0
QTi	QT interval	ms	240	230	230	230

- Notes:
1. A dash (—) in a SHAPE column means that this parameter is not specified.
 2. The amplitudes and durations are nominal values at LA-RA for an AMPLITUDE setting of 1 mV and a WIDTH setting of x2 using Fogg M310 & M311 ECG Simulators, or equivalent.
 3. A useful reference for ECG waveform definitions may be The American National Standard for Cardiac Monitors, Heart Rate Meters & Alarms, ANSI/AAMI, EC13-1992.